

, l 0, l 0oh dMyt] gki M+  
**S.S.V. COLLEGE, HAPUR**  
1/0kjh pj.k fl g fo"of0 | ky; e;B l s l Ec) ½



Date 20/11/2018

Tender Ref. No.- SSVV/DST-FIST/TF-01/2018/02

**Corrigendum**  
**For**  
**Tender for Supply and Installation of**  
**DOUBLE BEAM UV-VISIBLE SPECTROPHOTOMETER**

The following revised and additional specification will be added-

**Under the Technical specification section annexure (IA)**

**Point No. 7**

**For**

Spectral bandwidth (SBW)- fixed or variable bandwidth

**Read**

Spectral bandwidth (SBW)- fixed ( $\leq 1$ ) or variable bandwidth

**Point No. 8**

**For**

Stray Light- At 198 nm (KCl, TGA & BP/EP method)  $\leq 1$  %T

**Read**

Stray Light- At 198 nm (KCl)  $\leq 0.5$  %T

**Point No. 21**

**Accessories- Include D2/Xenon Lamp**

**Point No.-22**

**For**

Warranty and AMC- 3 Year warranty including all parts and source+ Two Years AMC/CMC quote separately.

**Read**

**Warranty and AMC- 2 Year warranty including all parts and source as well as Computer, Printer and UPS+ Two Years AMC quote Separately**

**Note-**

In the tender Documents where- 3 year warranty + two year AMC/CMC ( viz. Annexure( V) and Price Bid etc.) mentioned should be

**Read**

**2 year warranty including all parts & source as well as Computer, Printer and UPS + Two Years AMC quote Separately .**

**Principal/Coordinator**  
**DST-FIST**